



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/706,346	11/12/2003	Steven F. Bolling	ORQIS.007A	6824

20995 7590 05/29/2007
KNOBBE MARTENS OLSON & BEAR LLP
2040 MAIN STREET
FOURTEENTH FLOOR
IRVINE, CA 92614

EXAMINER

KOHARSKI, CHRISTOPHER

ART UNIT	PAPER NUMBER
----------	--------------

3763

NOTIFICATION DATE	DELIVERY MODE
-------------------	---------------

05/29/2007

ELECTRONIC

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

jcartee@kmob.com
eOAPilot@kmob.com

Office Action Summary

Application No.

10/706,346

Applicant(s)

BOLLING ET AL.

Examiner

Christopher D. Koharski

Art Unit

3763

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 16 January 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1,3-21 and 86-109 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☒ Claim(s) 7,10,11 and 99-103 is/are allowed.
- 6) ☒ Claim(s) 1,3-6,8,9,12-21,86-98 and 104-109 is/are rejected.
- 7) ☒ Claim(s) 106 is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☒ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☒ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date 1/16/07.
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- ☐ Notice of Informal Patent Application
- ☐ Other: _____.

DETAILED ACTION

Response to Amendment/Supplemental Action

Examiner acknowledges the reply filed 1/16/2007 in which claim 1 was amended, and new claims 107-109 added. Currently claims 1, 3-21 and 86-109 are pending for examination.

The following is a supplemental action to include dependant claims that were left out of the office action filed 3/26/2007.

Information Disclosure Statement

The information disclosure statement (IDS) that was submitted on 1/16/2007 is in compliance with the provisions of 37 CFR 1.97. Accordingly, the examiner is considering the information disclosure statement.

Allowable Subject Matter

Claims 10 and 11 are allowed.

Claims 7 and 99-103 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Claim Objections

Claim 106 is objected to because of the following informalities: Regarding claim 106, the claim improperly depends from itself. Appropriate correction is required.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

Art Unit: 3763

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 3-6, 8-9, 12-13, 18, 21, 86-98, 104 and 107-109 rejected under 35

U.S.C. 102(b) as being anticipated by Maria van Erp (6,102,891). Maria van Erp discloses a balloon catheter with valve.

Regarding claims 1, 3-6, 8-9, 12-13, 18, 21, 86-98, 104 and 107-109, Maria van Erp discloses a device (2) capable of being used as a percutaneous cannula comprising a main cannula portion with a blood flow lumen (5) (Figure 1) and tapered tip portion (closest exemplified by 4) near a distal end with discharge openings (14, 25) and a redirecting members (3, 24) (Figure 7) that are expandable under pressure of blood flow through the lumen, presenting a concave redirecting member (3, 24) that changes its perimeter and surface area over the discharge opening and is configured to discharge blood proximally along the cannula and is larger than the opening. The expandable members are elastic (3, 24) and capable of being collapsed to cover the openings during insertion and to be inflated to a predefined shape (circular, spherical) (Figure 1).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

Claims 14-17 and 19-20 are rejected under 35 U.S.C 103(a) as being unpatentable over Maria van Erp in view of Hedge et al. (6,231,543). Maria van Erp meets the claim limitations as described above except for a guidewire and seal in the distal end thereof.

However, Hedge et al. teaches a single lumen balloon catheter.

Regarding claims 14-17 and 19-20, Hedge et al. teaches a catheter that includes a guidewire (18) and a guidewire lumen (16) with multiple guidewire seals (30, 42, 44) that minimize blood flow through the lumen and can plug the lumen when a guidewire is not present in the lumen (Figures 4-6B).

At the time of the invention, it would have been obvious to add the guidewire and lumen of Hedge et al. to the system of Maria van Erp in order to aid in catheter guiding and tracking during a surgical procedure. The references are analogous in the art and with the instant invention; therefore, a combination is proper. Therefore, one skilled in the art would have combined the teachings in the references in light of the disclosure of Hedge et al.

Claim Rejections - 35 USC § 103

Claims 105 and 106 are rejected under 35 U.S.C 103(a) as being unpatentable over Maria van Erp.

The modified Maria van Erp discloses the claimed invention except for the flap being composed of a silicone with a material hardness of less than 15 measured on the A scale durometer. It would have been obvious to one having ordinary skill in the art at the time the invention was made to make the flaps out of silicone, since it has been held to be within the general skill of a worker in the art to select a known material on the basis of its suitability for the intended use as a matter of obvious design choice. *In re Leshin*, 227 F.2d 197, 125 USPQ 416 (CCPA 1960). In addition it would have been obvious to one having ordinary skill in the art at the time the invention was made to optimize the materials properties, since it has been held that discovering an optimum value of a result effective variable involves only routine skill in the art. *In re Boesch*, 617 F.2d 272, 205 USPQ 215 (CCPA 1980).

Response to Arguments

Applicant's arguments with respect to claims 1, 3-21 and 86-109 have been considered but are moot in view of the new ground(s) of rejection necessitated by applicant's amendments.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP

§ 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher D. Koharski whose telephone number is 571-272-7230. The examiner can normally be reached on 7:30am to 4:00pm EST.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nick Lucchesi can be reached on 571-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Date:

5/17/07

Christopher D. Koharski
AU 3763



NICHOLAS D. LUCCHESI
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700